

## Instructions for use

# SHBG ELISA

**REF****AA E-1200R****RUO**

For research  
use only –  
Not for use  
in diagnostic  
procedures

---

## SHBG ELISA

### 1. INTRODUCTION

#### 1.1 Intended Use

The **SHBG ELISA** is an enzyme immunoassay for the quantitative measurement of the sex-hormone-binding globulin (SHBG) in serum or plasma (EDTA-, heparin- or citrate plasma).

#### 1.2 Summary and Explanation

Sex-hormone-binding globulin (SHBG), a homodimeric glycoprotein of 95 kD, is synthesized in the liver and has a half-values time of 7 days in plasma. SHBG specifically binds steroid hormones with high affinity (DHT > testosterone > estrone/estradiol > DHEA/ androstenedione/ estriol), and its main function is sex-steroid transport within the blood stream and to extravascular target tissues. SHBG also plays a key role in regulating bioavailable sex-steroid concentrations through competition of sex steroids for available binding sites and fluctuations in SHBG concentrations. SHBG concentration in blood shows high inter-individual variability and is influenced by androgen/estrogen balance, nutritional status, body mass index, sex, insulin concentration among others.

### 2. PRINCIPLE OF THE TEST

The SHBG ELISA is a solid phase enzyme-linked immunosorbent assay (ELISA) based on the **sandwich principle**.

The microtiter wells are coated with a monoclonal (mouse) antibody directed towards a unique antigenic site of the SHBG molecule. An aliquot of sample containing endogenous SHBG is incubated in the coated well with enzyme conjugate, which is an anti-SHBG antibody conjugated with horseradish peroxidase. After incubation the unbound conjugate is washed off.

The amount of bound peroxidase conjugate is proportional to the concentration of SHBG in the sample.

Having added the substrate solution, the intensity of colour developed is proportional to the concentration of SHBG in the sample.


### 3. WARNINGS AND PRECAUTIONS

1. This kit is for research use only. For professional use only.
2. All reagents of this test kit which contain human serum or plasma have been tested and confirmed negative for HIV I/II, HBsAg and HCV by FDA approved procedures. All reagents, however, should be treated as potential biohazards in use and for disposal.
3. Before starting the assay, read the instructions completely and carefully. Use the valid version of instructions for use provided with the kit. Be sure that everything is understood.
4. The microplate contains snap-off strips. Unused wells must be stored at 2 °C to 8 °C in the sealed foil pouch and used in the frame provided.
5. Pipetting of samples and reagents must be done as quickly as possible and in the same sequence for each step.
6. Use reservoirs only for single reagents. This especially applies to the substrate reservoirs. Using a reservoir for dispensing a substrate solution that had previously been used for the conjugate solution may turn solution colored. Do not pour reagents back into vials as reagent contamination may occur.
7. Mix the contents of the microplate wells thoroughly to ensure good test results. Do not reuse microwells.
8. Do not let wells dry during assay; add reagents immediately after completing the rinsing steps.
9. Allow the reagents to reach room temperature (18 °C to 25 °C) before starting the test. Temperature will affect the absorbance readings of the assay. However, values for the samples will not be affected.
10. Never pipet by mouth and avoid contact of reagents and specimens with skin and mucous membranes.
11. Do not smoke, eat, drink or apply cosmetics in areas where specimens or kit reagents are handled.
12. Wear disposable latex gloves when handling specimens and reagents. Microbial contamination of reagents or specimens may give false results.
13. Handling should be done in accordance with the procedures defined by an appropriate national biohazard safety guideline or regulation.
14. Do not use reagents beyond expiry date as shown on the kit labels.
15. All indicated volumes have to be performed according to the protocol. Optimal test results are only obtained when using calibrated pipettes and microtiter plate readers.
16. Do not mix or use components from kits with different lot numbers. It is advised not to exchange wells of different plates even of the same lot. The kits may have been shipped or stored under different conditions and the binding characteristics of the plates may result slightly different.
17. Avoid contact with Stop Solution containing 0.5 M H<sub>2</sub>SO<sub>4</sub>. It may cause skin irritation and burns.










18. Some reagents contain Proclin 300, BND and/or MIT as preservatives. In case of contact with eyes or skin, flush immediately with water.
19. TMB substrate has an irritant effect on skin and mucosa. In case of possible contact, wash eyes with an abundant volume of water and skin with soap and abundant water. Wash contaminated objects before reusing them. If inhaled, take the person to open air.
20. Chemicals and prepared or used reagents have to be treated as hazardous waste according to the national biohazard safety guideline or regulation.
21. For information on hazardous substances included in the kit please refer to Safety Data Sheets. Safety Data Sheets for this product are available upon request directly from the manufacturer.

#### 4. REAGENTS

##### 4.1 Reagents provided


**AA E-1231**  **Microtiterwells**  
 Contents: 12 x 8 (break apart) strips, 96 wells;  
 Wells coated with anti-SHBG antibody (monoclonal).

**Standards and Controls** - ready to use

Cat. no.	Component	Standard	Concentration	Volume / Vial
AA E-1201	 STANDARD A	Standard A	0 nmol/l	0.5 ml
AA E-1202	 STANDARD B	Standard B	4 nmol/l	0.5 ml
AA E-1203	 STANDARD C	Standard C	16 nmol/l	0.5 ml
AA E-1204	 STANDARD D	Standard D	32 nmol/l	0.5 ml
AA E-1205	 STANDARD E	Standard E	65 nmol/l	0.5 ml
AA E-1206	 STANDARD F	Standard F	130 nmol/l	0.5 ml
AA E-1207	 STANDARD G	Standard G	260 nmol/l	0.5 ml
AA E-1251	 CONTROL 1	Low Control	For control values and ranges please refer to vial label or QC-Report.	0.5 ml
AA E-1252	 CONTROL 2	High Control		0.5 ml

*The standards are calibrated against the following reference material: WHO International Standard for Sex Hormone Binding Globulin (08/266)*

Contents: Contain preservative.

**AA E-1213**  **Assay Buffer** - ready to use

Contents: Contains preservative.

Volume: 1 x 125 ml

**AA E-1240**  **Enzyme Conjugate** - ready to use

Contents: Anti-SHBG antibody conjugated with horseradish peroxidase;  
 Contains preservative.

Volume: 1 x 14 ml

**AA E-1255**  **Substrate Solution** - ready to use

Contents: Tetramethylbenzidine (TMB).

Volume: 1 x 14 ml

**FR E-0080**  **Stop Solution** - ready to use

Contents: Contains 0.5 M H<sub>2</sub>SO<sub>4</sub>,  
 Avoid contact with the stop solution. It may cause skin irritations and burns.

Volume: 1 x 14 ml

Hazards identification:



H290 May be corrosive to metals.  
 H314 Causes severe skin burns and eye damage.

**FR E-0030**      **WASH-CONC 40x**      **Wash Solution - 40x concentrated**

Volume:            1 x 30 ml

See "Reagent Preparation".

#### **4.2 Materials required but not provided**

- A microtiter plate calibrated reader (450 ± 10 nm)
- Calibrated variable precision micropipettes.
- Absorbent paper.
- Distilled or deionized water
- Tubes for dilution of standards, controls and samples
- Timer
- Graph paper or software for data reduction

#### **4.3 Storage Conditions**

When stored at 2 °C - 8 °C unopened reagents will retain reactivity until expiration date. Do not use reagents beyond this date.

Opened reagents must be stored at 2 °C - 8 °C. Microtiter wells must be stored at 2 °C - 8 °C. Once the foil bag has been opened, care should be taken to close it tightly again.

Opened kits retain activity for 2 months if stored as described above.

#### **4.4 Reagent Preparation**

Bring all reagents and required number of strips to room temperature prior to use.

##### ***Wash Solution***

Add deionized water to the 40X concentrated Wash Solution.

Dilute 30 ml of concentrated *Wash Solution* with 1170 ml deionized water to a final volume of 1200 ml.

*The diluted Wash Solution is stable for 2 weeks at room temperature.*

#### **4.5 Disposal of the Kit**

The disposal of the kit must be made according to the national regulations. Special information for this product is given in the Safety Data Sheet, section 13.

#### **4.6 Damaged Test Kits**

In case of any severe damage to the test kit or components, the manufacturer has to be informed in writing, at the latest, one week after receiving the kit. Severely damaged single components should not be used for a test run. They have to be stored until a final solution has been found. After this, they should be disposed according to the official regulations.

### **5. SPECIMEN COLLECTION AND PREPARATION**

Serum or plasma (EDTA-, heparin- or citrate plasma) can be used in this assay. EDTA and citrate plasma samples may give slightly lower results.

*Please note:* Samples containing sodium azide should not be used in the assay.

#### **5.1 Specimen Collection**

##### **Serum:**

Collect blood by venipuncture (e.g. Sarstedt Monovette for serum), allow to clot, and separate serum by centrifugation at room temperature. Do not centrifuge before complete clotting has occurred.

##### **Plasma:**

Whole blood should be collected into centrifuge tubes containing anti-coagulant (e.g. Sarstedt Monovette with the appropriate plasma preparation) and centrifuged immediately after collection.

#### **5.2 Specimen Storage and Preparation**

Specimens should be capped and may be stored for up to 4 days at 2 °C - 8 °C prior to assaying.

Specimens held for a longer time (up to 3 months) should be frozen only once at -20 °C prior to assay.

Thawed samples should be inverted several times prior to testing.

### 5.3 Specimen Dilution

If in an initial assay, a specimen is found to contain more than the highest standard, the specimens can be further diluted with *Assay Buffer* and re-assayed as described in Assay Procedure. For the calculation of the concentrations this dilution factor has to be taken into account.

Example:

- a) dilution 1:10: 10 µl prediluted sample + 90 µl *Assay Buffer* (mix thoroughly)
- b) dilution 1:100: 10 µl dilution a) 1:10 + 90 µl *Assay Buffer* (mix thoroughly).

## 6. ASSAY PROCEDURE

### 6.1 General Remarks

- All reagents and specimens must be allowed to come to room temperature before use. All reagents must be mixed without foaming.
- Once the test has been started, all steps should be completed without interruption.
- Use new disposal plastic pipette tips for each standard, control or sample in order to avoid cross contamination.
- Absorbance is a function of the incubation time and temperature. Before starting the assay, it is recommended that all reagents are ready, caps removed, all needed wells secured in holder, etc. This will ensure equal elapsed time for each pipetting step without interruption.
- As a general rule the enzymatic reaction is linearly proportional to time and temperature.

### 6.2 Predilution of standards, controls and samples

Prior to the assay, all standards, controls and samples need to be diluted **1+100** in Assay Buffer

Example: 10 µl sample + 1000 µl *Assay Buffer*

**Thoroughly mix for 10 seconds.** It is important to have a complete mixing in this step.

Take **50 µl** of the prediluted standards, controls and samples for the SHBG ELISA

### 6.3 Test Procedure

Each run must include a standard curve.

1. Secure the desired number of Microtiter wells in the frame holder.
2. Dispense <b>50 µl</b> of each <u>prediluted</u> <b>Standard, Control</b> and <b>sample</b> with <u>new disposable tips</u> into appropriate wells.
3. Incubate for <b>120 minutes</b> at room temperature.
4. Briskly shake out the contents of the wells. Rinse the wells <b>3 times</b> with <b>300 µl - 400 µl</b> diluted <i>Wash Solution</i> per well. Strike the wells sharply on absorbent paper to remove residual droplets. <b>Important note:</b> The sensitivity and precision of this assay is markedly influenced by the correct performance of the washing procedure!
5. Dispense <b>100 µl Enzyme Conjugate</b> into each well.
6. Incubate for <b>30 minutes</b> at room temperature.
7. Briskly shake out the contents of the wells. Rinse the wells <b>3 times</b> with <b>300 µl - 400 µl</b> diluted <i>Wash Solution</i> per well. Strike the wells sharply on absorbent paper to remove residual droplets.
8. Add <b>100 µl of Substrate Solution</b> to each well.
9. Incubate for <b>15 minutes</b> at room temperature.
10. Stop the enzymatic reaction by adding <b>100 µl of Stop Solution</b> to each well.
11. Determine the absorbance (OD) of each well at <b>450 ± 10 nm</b> with a microtiter plate reader. It is recommended that the wells be read <b>within 10 minutes</b> after adding the <i>Stop Solution</i> .

## 6.4 Calculation of Results

1. Calculate the average absorbance values for each set of standards, controls and samples.
2. Using linear graph paper, construct a standard curve by plotting the mean absorbance obtained from each standard against its concentration with absorbance value on the vertical (Y) axis and concentration on the horizontal (X) axis.
3. Using the mean absorbance value for each sample determine the corresponding concentration from the standard curve.
4. Automated method: The results in the Instructions for Use have been calculated automatically using a 4-Parameter curve fit. (4 Parameter Rodbard or 4 Parameter Marquardt are the preferred methods.) Other data reduction functions may give slightly different results.
5. The concentration of the samples can be read directly from this standard curve. Samples with concentrations higher than that of the highest standard have to be further diluted or reported as > 260 nmol/l. For the calculation of the concentrations this dilution factor has to be taken into account.

### 6.4.1 Example of Typical Standard Curve

The following data is for demonstration only and cannot be used in place of data generations at the time of assay.

Standard		Optical Units (450 nm)
Standard A	0 nmol/l	0.02
Standard B	4 nmol/l	0.09
Standard C	16 nmol/l	0.27
Standard D	32 nmol/l	0.49
Standard E	65 nmol/l	0.84
Standard F	130 nmol/l	1.36
Standard G	260 nmol/l	1.93

## 7. EXPECTED NORMAL VALUES

It is strongly recommended that each laboratory should determine its own normal and abnormal values.

In a study conducted with apparently healthy subjects, using the SHBG ELISA the following data were observed:

Population	n	Mean (nmol/l)	Median (nmol/l)	2.5 <sup>th</sup> - 97.5 <sup>th</sup> Percentile (nmol/l)	Range (min - max) (nmol/l)
Males	78	45.3	42.1	17.7 - 92.8	16.8 - 113.2
Females	40	65.0	58.2	20.4 - 126.7	16.1 - 128.4

## 8. QUALITY CONTROL

Good laboratory practice requires that controls be run with each standard curve. A statistically significant number of controls should be assayed to establish mean values and acceptable ranges to assure proper performance.

It is recommended to use control samples according to state and federal regulations. The use of control samples is advised to assure the day to day validity of results. Use controls at both normal and pathological levels.

The controls and the corresponding results of the QC-Laboratory are stated in the QC certificate added to the kit. The values and ranges stated on the QC sheet always refer to the current kit lot and should be used for direct comparison of the results.

It is also recommended to make use of national or international Quality Assessment programs in order to ensure the accuracy of the results.

Employ appropriate statistical methods for analysing control values and trends. If the results of the assay do not fit to the established acceptable ranges of control materials sample results should be considered invalid.

In this case, please check the following technical areas: Pipetting and timing devices; photometer, expiration dates of reagents, storage and incubation conditions, aspiration and washing methods.

After checking the above mentioned items without finding any error contact your distributor or the manufacturer directly.

## 9. PERFORMANCE CHARACTERISTICS

### 9.1 Assay Dynamic Range

The range of the assay is between 0.408 - 260 nmol/l.

### 9.2 Specificity of Antibodies (Cross Reactivity)

The following substances were tested for cross reactivity of the assay:

Substance	% Cross-reactivity
Corticoid binding globulin	< 0.2
Thyroxin binding globulin	< 0.04

### 9.3 Sensitivity

The analytical sensitivity of the ELISA was calculated by adding 2 standard deviations to the mean of 20 replicate analyses of the *Standard A* and was found to be 0.23 nmol/l.

The Limit of Blank (LoB) is 0.23 nmol/l.

The Limit of Detection (LoD) is 0.408 nmol/l.

The Limit of Quantification (LoQ) is 0.757 nmol/l.

### 9.4 Reproducibility

#### 9.4.1 Intra Assay

The within assay variability is shown below:

Sample	n	Mean (nmol/l)	CV (%)
1	10	41.67	2.3
2	10	66.75	4.6
3	10	87.37	3.2
4	10	133.62	4.8

#### 9.4.2 Inter Assay

The between assay variability is shown below:

Sample	n	Mean (nmol/l)	CV (%)
1	30	41.99	5.7
2	30	68.94	6.3
3	30	90.00	6.2
4	30	136.96	5.2

#### 9.4.3 Inter-Lot

The inter-assay (between-lots) variation was determined by repeated measurements of samples with 3 different kit lots.

Sample	n	Mean (nmol/l)	CV (%)
1	18	44.04	8.1
2	18	61.32	8.9
3	18	92.27	11.7
4	18	160.92	8.2

### 9.5 Recovery

Recovery of the ELISA was determined by adding increasing amounts of the analyte to different samples containing different amounts of endogenous analyte.

	Sample 1	Sample 2	Sample 3	Sample 4	
Concentration (nmol/l)	45.5	76.8	85.62	158.6	
Average Recovery (%)	95.9	92.6	86.7	89.2	
Range of Recovery (%)	from	92.9	88.3	85.5	87.4
	to	99.8	96.9	87.7	90.5

## 9.6 Linearity

	Sample 1	Sample 2	Sample 3	Sample 4	
<b>Concentration (nmol/l)</b>	44.5	73.4	98.5	177.6	
<b>Average Recovery (%)</b>	98.6	97.3	98.5	99.2	
<b>Range of Recovery (%)</b>	from	96.1	93.8	94.2	96.2
	to	101.0	100.1	100.4	101.6

## 10 LIMITATIONS OF USE

Reliable and reproducible results will be obtained when the assay procedure is performed with a complete understanding of the package insert instruction and with adherence to good laboratory practice. Any improper handling of samples or modification of this test might influence the results.

### 10.1 Interfering Substances

Haemoglobin (up to 4 mg/ml), Bilirubin (up to 0.5 mg/ml) and Triglyceride (up to 7.5 mg/ml) have no influence on the assay results.

### 10.2 Drug Interferences

Until today no substances (drugs) are known to us, which have an influence to the measurement of SHBG in a sample.

### 10.3 High-Dose-Hook Effect







Hook effect was not observed in this test up to a concentration of 11350 nmol/l of SHBG.



**11. REFERENCES / LITERATURE**

1. Moore, JW and Bulbrook RD. The epidemiology and function of sex hormone binding globulin. IN Oxford Reviews of Reproductive Biology, 1988; 10: 180 - 236.
2. Selby, C. Sex hormone binding globulin: origin, function and clinical significance. Ann. Clin. Biochem. 1990; 27: 532 - 541.
3. Tehernof A, Despres JP: Sex steroid hormone, sex hormone-binding globulin, and obesity in men and women. Horm. Metab. Res.2000; 32:526-536.
4. Kahn SM, Hryb DJ, Nakhle AM, Romas NA: Sex hormone-binding globulin is synthesized in target cells. J Endocrinol 2002; 175:113-120.
5. Hammond GL: Access of reproductive steroids to target issues. Obstet Gynecol Clin North Am 2002; 29:411-423.
6. 5. Elmlinger MW, Kuhnel W, Ranke MB: Reference ranges for serum concentrations of lutropin (LH), follitropin (FSH), estradiol (E2), prolactin, progesterone, sex hormone binding globulin (SHBG), dehydroepiandrosterone sulfate (DHEA-S), cortisol and ferritin in neonates, children, and young adults. Clin Chem Lab Med 2002; 40(11):1151-1160.
7. Deswal ., Yadav A, Dang AS Sex hormone binding globulin - an important biomarker for predicting PCOS risk: A systematic review and meta-analysis. Syst Biol Reprod Med 2018; 64: 12-24.
8. Sanches de Melo et al. Hormonal contraception in women with polycystic ovary syndrome: choices, challenges, and noncontraceptive benefits. Open Access J Contracept.2017; 2;8:13-23.
9. Pugeat et al. (2018) Hyperandrogenic states in women: pitfalls in laboratory diagnosis. Eur J Endocrin 2018; 178, 141-54.
10. Rothman MS, Wierman ME. How should postmenopausal androgen excess be evaluated? Clin Endocrinol 2011; 75(2):160-4.
11. Wu FCW et al. Identification of Late-Onset Hypogonadism in Middle-Aged and Elderly Men. N Engl J Med 2010; 363:123-35.
12. Tajar A et al. Characteristics of secondary, primary, and compensated hypogonadism in aging men: evidence from the European Male Ageing Study. J Clin Endocrinol Metab 2010 95(4):1810-8.

**Symbols:**

	Storage temperature		Manufacturer		Contains sufficient for <n> tests
	Expiry date	<b>LOT</b>	Batch code		
	Consult instructions for use	<b>CONT</b>	Content		
	Caution	<b>REF</b>	Catalogue number	<b>RUO</b>	For research use only!