

CERTIFICATE OF ANALYSIS

Product Name:	Prolex™ Stre	eptococcal Grouping Latex Kit	Lot Number:	C18806
Product Code:	PL.030SJ		Expiration Date:	2022-01-31
Product Code	Quantity	Components	•	
PL.031S	1 x 3.0 ml	Streptococcal Group A Lat	tex Reagent	
PL.032S	1 x 3.0 ml	Streptococcal Group B Lat	tex Reagent	
PL.033S	1 x 3.0 ml	Streptococcal Group C La	reptococcal Group C Latex Reagent	
PL.034S	1 x 3.0 ml	Streptococcal Group D La	treptococcal Group D Latex Reagent	
PL.035S	1 x 3.0 ml	Streptococcal Group F Lat	tex Reagent	
PL.036S	1 x 3.0 ml	Streptococcal Group G La	tex Reagent	
PL.037	1 x 3.2 ml	Streptococcal Extraction R	Reagent 1	
PL.038	1 x 3.2 ml	Streptococcal Extraction R	Reagent 2	
PL.039	2 x 8.0 ml	Streptococcal Extraction R	Reagent 3	
PL.040S	1 x 2.0 ml	Polyvalent Positive Contro	I (A,B,C,D,F,G)	
	300	Mixing Sticks		
PL.092-48J	48 cards	Test Cards		
Additional Information				
Storage Conditions: 2-8°C				
Precautions:		Refer to Instructions for Use		
Quality Control Testing				
Criteria				Result
Chemical extraction of Strep group specific antigens.				PASS
				5400
Strong agglutination of the test isolate when tested against the corresponding latex reagent. PAS				PASS
No agglutination reaction of the test isolate when tested against heterologous latex reagents.				PASS
Strong agglutination of the Positive Control when tested against the latex reagents. PASS				PASS
Release				
			S. A.	
	2020 08 17		Syanlam	_
Date			Quality Control	

Canada 🕾 800 268 2341 Fax 905 731 0206 20 Mural Street, Unit #4, Richmond Hill, Ontario, L4B 1K3 support@pro-lab.com U.S.A. 🕾 800 522 7740 Fax 800 332 0450 21 Cypress Blvd., Suite 1155, Round Rock, Texas 78665-1046 support@pro-lab.us U.K. 🕾 0151 353 1613 Fax 0151 353 1614 3 Bassendale Road, Bromborough, Wirral, Merseyside CH62 3QL uksupport@pro-lab.com