

### CERTIFICATE OF ANALYSIS

|  |                                    |
|--|------------------------------------|
| <b>Product Name:</b> Streptococcal Grouping Reagent Latex Suspension Group F | <b>Lot Number:</b> C18062B         |
| <b>Product Code:</b> PL.035S   | <b>Expiration Date:</b> 2021-05-31 |


#### Additional Information

|                     |                               |
|---------------------|-------------------------------|
| Storage Conditions: | 2-8°C                         |
| Precautions:        | Refer to Instructions for Use |

#### Quality Control Testing

| Appearance  | Reaction | Result       |
|---|----------|--------------|
| Homogenous suspension with no agglutination           | N/A      | Satisfactory |
| <b>Organism</b>                                       |          |              |
| <i>Streptococcus</i> sp. ATCC 12392 (Group F, type 2) | Positive | PASS         |
| <i>Streptococcus</i> Group F - Clinical Isolate       | Positive | PASS         |
| <i>Streptococcus pyogenes</i> ATCC 19615 (Group A)    | Negative | PASS         |
| <i>Streptococcus agalactiae</i> ATCC 12386 (Group B)  | Negative | PASS         |
| <i>Streptococcus equisimilis</i> ATCC 12388 (Group C) | Negative | PASS         |
| <i>Enterococcus faecalis</i> ATCC 49477 (Group D)     | Negative | PASS         |
| <i>Streptococcus</i> sp. ATCC 12394 (Group G)         | Negative | PASS         |
| <i>Listeria monocytogenes</i> NCTC 11994              | Negative | PASS         |
| <b>Control</b>  |          |              |
| Polyvalent Positive Control (PL.040/PL.040S)          | Positive | PASS         |

#### Release

|            |  |
|------------|--|
| 2019 12 09 |  |
| Date       | Quality Control  |

Canada ☎ 800 268 2341 Fax 905 731 0206  
 20 Mural Street, Unit #4, Richmond Hill, Ontario,  
 L4B 1K3  
[support@pro-lab.com](mailto:support@pro-lab.com)

U.S.A. ☎ 800 522 7740 Fax 800 332 0450  
 21 Cypress Blvd., Suite 1155, Round Rock,  
 Texas 78665-1046  
[support@pro-lab.us](mailto:support@pro-lab.us)

U.K. ☎ 0151 353 1613 Fax 0151 353 1614  
 3 Bassendale Road, Bromborough, Wirral,  
 Merseyside CH62 3QL  
[uksupport@pro-lab.com](mailto:uksupport@pro-lab.com)