



www.pro-lab.com

CERTIFICATE OF ANALYSIS

Product Name: Microbank™ – Mixed		Lot Number: C18642M
Product Code: PL.170/M		Expiration Date: 2022-05-31
Product Code	Quantity	Components
PL.170/M	80	Cryovials containing approximately 25 beads and 1 ml of cryopreservative
	1	Retrieval device
	1	Numbered storage container
Additional Information		
Storage Conditions:		15-30°C
Precautions:		Refer to Instructions for Use
Quality Control Testing		
Sterility	Criteria	Result
	Conforms	PASS
Visual Inspection	Conforms	PASS
Organism	Recovery of Organism	Result
<i>Escherichia coli</i> ATCC 25922	Satisfactory	PASS
<i>Staphylococcus aureus</i> ATCC 25923	Satisfactory	PASS
<i>Klebsiella pneumoniae</i> ATCC 13883	Satisfactory	PASS
<i>Clostridium perfringens</i> ATCC 13124	Satisfactory	PASS
<i>Neisseria gonorrhoeae</i> ATCC 43069	Satisfactory	PASS
Release		
2020 06 22		
Date	Quality Control	

Canada ☎ 800 268 2341 Fax 905 731 0206
20 Mural Street, Unit #4, Richmond Hill, Ontario,
L4B 1K3
support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450
21 Cypress Blvd., Suite 1070, Round Rock,
Texas 78665-1034
ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
3 Bassendale Road, Bromborough, Wirral,
Merseyside CH62 3QL
uksupport@pro-lab.com