LABOR DIAGNOSTIKA NORD GmbH & Co.KG | Am Eichenhain 1 | 48531 Nordhorn | Germany | Tel. +49 5921 8197-0 | Fax +49 5921 8197-222 | info@ldn.de | www.ldn.de

# Instructions for use FT3 ELISA 2nd Generation











# fT3 ELISA 2nd Generation

#### INTENDED USE

For the direct quantitative determination of Free Triiodothyronine by enzyme immunoassay in human serum. For *in vitro* diagnostic use only.

#### PRINCIPLE OF THE TEST

The principle of the following enzyme immunoassay test follows the typical competitive binding scenario. Competition occurs between an unlabeled antigen (present in standards, controls and patient samples) and an enzyme-labelled antigen (conjugate) for a limited number of antibody binding sites on the microwell plate. The washing and decanting procedures remove unbound materials. After the washing step, the enzyme substrate is added. The enzymatic reaction is terminated by addition of the stopping solution. The absorbance is measured on a microtiter plate reader. The intensity of the colour formed is inversely proportional to the concentration of fT3 in the sample. A set of standards is used to plot a standard curve from which the amount of fT3 in patient samples and controls can be directly read.

The labelled T3 (conjugate) employed in this assay system has shown no substantial binding properties towards thyroxine-binding globulin (TBG) or human serum albumin (HSA). The binding sites on the microwell plates are designed to be of a low binding-capacity in order not to disturb the equilibrium between T3 and its carrying proteins. The assay is carried out under normal physiological conditions of pH, temperature and ionic strength.

### **CLINICAL APPLICATIONS**

Triiodothyronine (T3) is a thyroid hormone found circulating in the bloodstream. T3 contains three iodine atoms and is produced largely through the extrathyroidal conversion of thyroxine (T4), the principal thyroid hormone with four iodine atoms. Most of the T3 that circulates in the blood is bound to carrier proteins such as TBG, prealbumin and albumin. The free fraction of T3 (fT3), which represents only 0.25% of the total amount, is considered to be the physiological active fraction.

Total T3 levels depend not only on thyroid status and the peripheral conversion of T4 to T3, but also on the concentration of thyroid hormone-binding proteins. Free T3 (fT3) on the other hand, is largely unaffected by variations in these carrier proteins which can occur under conditions such as pregnancy, estrogen therapy and the use of oral contraceptives. Therefore, free T3 typically reflects a patient's actual thyroid status more reliably than total T3.

Measurement of free T3 is generally recommended for patients with symptoms of hyperthyroidism as found in Graves' disease, toxic adenoma and toxic multinodular goiter.

# PROCEDURAL CAUTIONS AND WARNINGS

- Users should have a thorough understanding of this protocol for the successful use of this kit. Reliable performance will only be attained by strict and careful adherence to the instructions provided.
- Control materials or serum pools should be included in every run at a high and low level for assessing the reliability of results.
- When the use of water is specified for dilution or reconstitution, use deionized or distilled water.
- In order to reduce exposure to potentially harmful substances, gloves should be worn when handling kit reagents and human specimens.
- All kit reagents and specimens should be brought to room temperature and mixed gently but thoroughly before use. Avoid repeated freezing and thawing of reagents and specimens.
- A calibrator curve must be established for every run.
- The controls should be included in every run and fall within established confidence limits.
- Improper procedural techniques, imprecise pipetting, incomplete washing as well as improper reagent storage may be indicated when assay values for the controls do not reflect established ranges.
- When reading the microplate, the presence of bubbles in the microwells will affect the optical densities (ODs). Carefully remove any bubbles before performing the reading step.
- The substrate solution (TMB) is sensitive to light and should remain colourless if properly stored. Instability or contamination may be indicated by the development of a blue colour, in which case it should not be used.
- When dispensing the substrate and stopping solution, do not use pipettes in which these liquids will come into contact with any metal parts.
- To prevent contamination of reagents, use a new disposable pipette tip for dispensing each reagent, sample, standard and control.
- Do not mix various lot numbers of kit components within a test and do not use any component beyond the expiration date printed on the label.
- Kit reagents must be regarded as hazardous waste and disposed of according to national regulations.

Version: 4.1 Effective: August 09, 2013 2/8

#### **LIMITATIONS**

- All the reagents within the kit are calibrated for the direct determination of fT3 in human serum. The kit is not calibrated for the determination of fT3 in other specimens of human or animal origin.
- Do not use grossly hemolyzed, grossly lipemic, icteric or improperly stored serum.
- Any samples or control sera containing azide or thimerosal are not compatible with this kit, as they may lead to false results.
- Samples reading higher than 40 pg/ml should be reported as such and should not be diluted. Dilution will alter the existing equilibrium and may lead to false results.
- The interpretation of free T3 results can be complicated by a variety of drugs, severe nonthyroidal illness and some rare conditions such as familial dysalbuminemic hyperthyroxinemia (FDH). For diagnostic purposes, the results of this assay should always be used in combination with the clinical examination, medical history and other findings.

## SAFETY CAUTIONS AND WARNINGS POTENTIAL BIOHAZARDOUS MATERIAL

Human serum that may be used in the preparation of the standards and controls has been tested and found to be non-reactive for Hepatitis B surface antigen and has also been tested for the presence of antibodies to HCV and Human Immunodeficiency Virus (HIV) and found to be negative. However no test method can offer complete assurance that HIV, HCV and Hepatitis B virus or any infectious agents are absent. The reagents should be considered a potential biohazard and handled with the same precautions as applied to any blood specimen.

#### CHEMICAL HAZARDS

Avoid contact with reagents containing TMB, hydrogen peroxide and sulfuric acid. If contacted with any of these reagents, wash with plenty of water. TMB is a suspected carcinogen.

## SPECIMEN COLLECTION AND STORAGE

Approximately 0.1 ml of serum is required per duplicate determination. Collect 4-5 ml of blood into an appropriately labelled tube and allow it to clot. Centrifuge and carefully remove the serum layer. Store at  $4^{\circ}$ C for up to 24 hours or at  $-10^{\circ}$ C or lower if the analyses are to be done at a later date. Consider all human specimens as possible biohazardous materials and take appropriate precautions when handling.

#### SPECIMEN PRETREATMENT

This assay is a direct system; no specimen pretreatment is necessary.

## REAGENTS AND EQUIPMENT NEEDED BUT NOT PROVIDED

- Precision pipettes to dispense 25, 50, 100, 150 and 300 μl
- Disposable pipette tips
- Distilled or deionized water
- A 37 °C incubator
- Microwell plate reader with a filter set at 450nm and an upper OD limit of 3.0 or greater\* (see assay procedure step 10).

#### REAGENTS PROVIDED

AA E-0030 WASH-CONC 10x Wash Buffer Concentrate - X10

Contents: One bottle containing buffer with a non-ionic detergent and a non-mercury preservative.

Volume: 50 mL/bottle

Storage: Refrigerate at 2-8°C

Stability: 12 months or as indicated on label.

Preparation: Dilute 1:10 in distilled or deionized water before use. If the whole plate is to be used dilute

50 mL of the wash buffer concentrate in 450 mL of water.

AA E-0055 SUBSTRATE TMB Substrate - Ready To Use.

Contents: One bottle containing tetramethylbenzidine and hydrogen peroxide in a non-DMF or DMSO

containing buffer.

Volume: 16 mL/bottle

Storage: Refrigerate at 2-8°C

Stability: 12 months or as indicated on label.

Version: 4.1 Effective: August 09, 2013 3/8

AA E-0080 STOP-SOLN Stopping Solution - Ready To Use.

Contents: One vial containing 1M sulfuric acid.

Volume: 6 mL/bottle

Storage: Refrigerate at 2-8°C

Stability: 12 months or as indicated on label.

#### Calibrators and Controls- Ready To Use.

Listed below are approximate concentrations, please refer to vial labels for exact concentrations:

Cat. no.	Symbol	Calibrator	Concentration	Volume/Vial
TF E-2101	STANDARD A	Calibrator A	0 pg/ml	0.5 ml
TF E-2102	STANDARD B	Calibrator B	1.7 pg/ml	0.5 ml
TF E-2103	STANDARD C	Calibrator C	3.2 pg/ml	0.5 ml
TF E-2104	STANDARD D	Calibrator D	5.8 pg/ml	0.5 ml
TF E-2105	STANDARD E	Calibrator E	13.5 pg/ml	0.5 ml
TF E-2106	STANDARD F	Calibrator F	40 pg/ml	0.5 ml
TF E-2151	CONTROL 1	Control 1	Refer to vial labels for expected	0.5 ml
TF E-2152	CONTROL 2	Control 2	value and acceptable range!	0.5 ml

Contents: fT3 in a human serum-based matrix with a non-mercury preservative. Prepared by spiking

serum with a defined quantity of T3.

Storage: Refrigerate at 2-8°C

Stability: 12 months in unopened vials or as indicated on label. Once opened, the standards should be

used within 14 days or aliquoted and stored frozen. Avoid multiple freezing and thawing

cycles.

TF E-2113 ASSAY-BUFF Assay Buffer - Ready To Use.

Contents: One vial containing a protein-based buffer with a non-mercury preservative.

Volume: 15 mL/bottle

Storage: Refrigerate at 2-8°C

Stability: 12 months or as indicated on label.

Ready To Use.

Contents: One 96 well (12x8) polyclonal antibody-coated microwell plate in a resealable pouch with

desiccant.

Storage: Refrigerate at 2-8°C

Stability: 12 months or as indicated on label.

TF E-2140 CONJUGATE-CONC 50x fT3-Horseradish Peroxidase (HRP) Conjugate Concentrate - X50

Contents: fT3-HRP conjugate in a protein-based buffer with a non-mercury preservative.

Volume: 0.3 mL/vial

Storage: Refrigerate at 2-8°C

Stability: 12 months or as indicated on label.

Preparation: Dilute 1:50 in assay buffer before use (eg. 40 µL of HRP in 2 mL of assay buffer). If the

whole plate is to be used dilute 240  $\mu L$  of HRP in 12 mL of assay buffer. Discard any that is

left over.

Version: 4.1 Effective: August 09, 2013 4/8

#### ASSAY PROCEDURE

All reagents must reach room temperature before use. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.

- 1. Prepare working solutions of the anti-fT3-HRP and wash buffer.
- **2.** Remove the required number of microwell strips. Reseal the bag and return any unused strips to the refrigerator.
- **3. Pipette 25 μI** of each **calibrator, control and specimen sample** into correspondingly labelled wells in duplicate.
- **4.** Pipette **100** μ**L** of the **conjugate working solution** into each well.
- 5. Gently shake the plate for 10 seconds. **Incubate** the plate at **37°C** for **1 hour**.
- 6. Wash the wells **3 times** with prepared wash buffer (**300 μL**/well for each wash) and tap the plate firmly against absorbent paper to ensure that it is dry (The use of a washer is recommended).
- 7. Pipette **150** µL of **TMB substrate** into each well at timed intervals.
- **8.** Incubate the plate at **37°C** for **10-15** minutes (or until Calibrator A attains dark blue colour for desired OD).
- **9.** Pipette **50**  $\mu$ I of **stopping solution** into each well at the same timed intervals as in step 7.
- **10.** Read the plate on a microwell plate reader at **450 nm** within 20 minutes after addition of the stopping solution.
- If the OD exceeds the upper limit of detection or if a 450nm filter is unavailable, a 405 or 415nm filter may be substituted. The optical densities will be lower, however, this will not affect the results of patient/control samples.

### **CALCULATIONS**


- Calculate the mean optical density of each calibrator duplicate.
- Draw a calibrator curve on semi-log paper with the mean optical densities on the Y-axis and the calibrator concentrations on the X-axis. If immunoassay software is being used, a 4-parameter or 5-parameter curve is recommended.
- Calculate the mean optical density of each unknown duplicate.
- Read the values of the unknowns directly off the calibrator curve.

## TYPICAL TABULATED DATA

Calibrator	OD :	OD 2	Mean OD	Value (pg/ml)
Α	2.559	2.548	2.554	0
В	2.401	2.379	2.390	1.7
С	2.171	2.146	2.159	3.2
D	1.564	1.568	1.566	5.8
Е	0.858	0.849	0.853	13.5
F	0.491	0.500	0.496	40
Unknown	1.868	1.872	1.870	4.4

## TYPICAL CALIBRATOR CURVE

Sample curve only. **Do not** use to calculate results.



Version: 4.1 Effective: August 09, 2013 5/8

### PERFORMANCE CHARACTERISTICS

#### **SENSITIVITY**

The lower detection limit is calculated from the standard curve by determining the resulting concentration of the mean OD of Calibrator A (based on 10 replicate analyses) minus 2 SD. Therefore, the sensitivity of the Direct fT3 ELISA kit is

0.3 pg/ml.

# SPECIFICITY (CROSS REACTIVITY)

The following compounds were tested for cross-reactivity with the Direct fT3 ELISA kit with T3 cross-reacting at 100%.

Compound	%Cross Reactivity
L-Triiodothyronine	100
D-Triiodothyronine	34
Triiodothyropropionic	20
acid	
Diiodo-D-thyronine	0.5
D-Thyroxine	0.3
L-Thyroxine	0.9

The following compounds were tested but cross-reacted at less than 0.1%: Diiodotyrosine, Iodotyrosine, Phenytoin,

Sodium Salicylate and r-Triiodothyronine.

### INTRA-ASSAY PRECISION

Three samples were assayed ten times each on the same calibrator curve. The results (in pg/ml) are tabulated below:

Sample Mean		SD	CV%
1	5.182	0.501	9.7
2	8.560	0.598	7.0
3	48.200	1.686	3.5

# INTER-ASSAY PRECISION

Three samples were assayed ten times over a period of four weeks. The results (in pg/ml) are tabulated below:

Sample	Mean	SD	CV%
1	3.306	0.284	8.6
2	5.154	0.402	7.8
3	8.698	0.713	8.2

## **EXPECTED NORMAL VALUES**

As for all clinical assays each laboratory should collect data and establish their own range of expected normal values. The following reference range (pg/ml) was established with 44 apparently healthy adults:

Ĭ	Group	N	Mea n	Central 95% Range
	Euthyro id Adults	44	3.7	2.2 - 5.3

Version: 4.1 Effective: August 09, 2013 6/8

## EFFECT OF THYROXINE BINDING GLOBULIN (TBG)

The purpose of this study was to investigate a possible interference caused by the binding of TBG to the fT3-HRP conjugate. The zero calibrator was spiked with purified TBG and assayed. The results are tabulated below:

TBG (µg/ml added)	OD	% B/B <sub>0</sub>
0	1.25 5	100
12.5	1.22 9	98
25	1.17 0	93
50	1.13 7	91
100	1.16 8	93
200	1.17 4	94
400	1.11 8	89

The results show no binding of labeled T3 to TBG even at higher than normal levels. In conclusion, results showed that there was no significant influence by TBG in the Direct Free T3 Direct ELISA Kit.

# EFFECT OF HUMAN SERUM ALBUMIN (HSA)

The purpose of this study was to investigate a possible interference of HSA on the assay procedure. The zero calibrator was spiked with purified HSA and assayed. The results are tabulated below:

HSA (mg/ml added)	OD	% B/B <sub>0</sub>
0	1.255	100
3.125	1.228	98
6.25	1.331	100
12.5	1.245	99
25	1.197	95
50	1.217	97
100	1.063	85

The results show no significant binding of labeled T3 to HSA even at higher than normal levels.

# EFFECT OF NON-ESTERIFIED FATTY ACIDS (NEFA)

The purpose of this study was to investigate a possible interference of NEFA on the assay procedure. Two samples were spiked with oleic acid and assayed. The results are tabulated below:

NEFA (mmol/L added)	Sample 1 (pg/ml)	Sample 2 (pg/ml)
0	4.4	8.7
0.5	4.6	7.5
3.5	4.6	8.3
25	4.6	10.9

The results show that NEFA may increase the free T3 values, only at higher than normal concentrations.

#### **EFFECT OF LIPEMIA**

The purpose of this study was to investigate a possible interference of lipemic samples on the assay procedure. Two samples were spiked with triglycerides and assayed. The results are tabulated below:

Triglycerid es (mg/dl added)	Sample1 (pg/ml)	Sample2 (pg/ml)	
0	4.4	8.7	
50	5.5	9.9	
75	5.9	10.8	

Results show that lipemic samples may increase the free T3 values. Therefore, lipemic samples should not be used in this assay.

Version: 4.1 Effective: August 09, 2013 7/8

#### REFERENCES

- Mullinger AN, et al. Free triiodothyronine in nonthyroidal illness (Letter). Clin. Chem. 1993;39:1555.
- Wilkins TA, et al. Assay performance and tracer properties for two analog-based of free triiodothyronine, Clin. Chem. 1986;32:465.
- Ekins RP. Measurement of free hormones in blood (Review), Endocri. Rev. 1990;11:5.
- John R, et al, Concentration of free thyroxine and free triiodothyronine in serum with patients with thyronine and triiodothyronine binding autoantibodies. Clin. Chem. 1990;36:470.
- Chopra IJ. RIA of iodothyronines. In: Abraham GE eds. Handbook of radioimmunoassay. New York: Marcel Dekker Inc, 1997:679.
- Demers LM. Thyroid function testing and automation. J.Clin. Ligand Assay. 1999;22:38.
- Kalra J, et al. Value of free thyroxine (FT4), free triiodothyronine (FT3) and sensitive thyrotropin (TSH) assay in the assessment of optimal thyronine therapy. Clin.Biochem. 1987;20:265.
- Bergmann PJ, et al. Free triiodothyronine in hypothyroidism and nonthyroidal illness (letter). Clin. Chem. 1994;40:4996.
- Cohen JH, et al. Thyrotoxicosis due to ingestion of excess thyroid hormone. Endocri. Rev. 1993;10:113.
- Ingbar SH, et al. A new method for measuring the free thyroid hormone in human serum and an analysis of the factors that interference its concentration. J. Clin. Invest. 1965;44:1679.
- Pedersen KO. Simultaneous determination of the free thyroxine and triiodothyronine fractions in serum. Scan. J. Clin. Lab. Invest. 1974;34:241.
- Weeke J, et al. Ultrasensitive radioimmunoassay for direct determination of free triiodothyronine concentration in human serum. Scan. J. Clin. Lab. Invest. 1975;35:237.
- Oddie TH et al. Triiodothyronine turnover in euthyroid subjects. J. Clin. Endocr.1971;33:653.
- Nauman JA et al. Total and free triiodothyronine in human serum. J. Clin. Invest. 1967;46:1346.

## Symbols:

+ <u>2</u> /	Storage temperature	***	Manufacturer	Σ	Contains sufficient for <n> tests</n>
	Expiry date	LOT	Batch code	I V D	For in-vitro diagnostic use only!
[]i	Consult instructions for use	CONT	Content	CE	CE labelled
Â	Caution	REF	Catalogue number	RUO	For research use only!

Version: 4.1 Effective: August 09, 2013 8/8